

Biomechanic Comparison of the Teno Fix Tendon Repair Device With the Cruciate and Modified Kessler..

Last Updated Wednesday, 28 March 2007

Journal of Hand Surgery

Scott W. Wolfe MD, et al. - Conclusions: The Teno Fix repair system did not confer a mechanical advantage over the locked cruciate or modified Kessler suture techniques for zone II lacerations in cadaveric flexor tendons during cyclic loading in a linear testing model. This information may help to define safe boundaries for postoperative rehabilitation when using this internal tendon repair device

Journal of Hand Surgery

Scott W. Wolfe MD, et al. - Conclusions: The Teno Fix repair system did not confer a mechanical advantage over the locked cruciate or modified Kessler suture techniques for zone II lacerations in cadaveric flexor tendons during cyclic loading in a linear testing model. This information may help to define safe boundaries for postoperative rehabilitation when using this internal tendon repair device

Read article: <http://www.mdlinx.com/OrthoLinx/xml-article.cfm/1793684>.